



मुख्यमंत्री निःशुल्क जांच यो

Rajasthan Medical Services Corporation Ltd.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN:U24232RJ2011SGC035067

E-Mail -rMSC@nic.in, edepmrmSC-ri@nic.in

No. F-8(215) RMSC/EPM/M-4/NIB-579/2020-21/ 1095

Date: 17/09/2021

Office Order


List of bidders declared responsive/non responsive as per documentary & technical evaluation in bid number No. F-8(215) RMSC/EPM/M-4/NIB-579/2020-21/510 Dated 22.03.2021 for item TRANSCUTANEOUS BILIRUBINOMETER 2.RADIANT WARMER 3. RADIANT WARMER WITH OXYGEN-PROVISION 4. LED PHOTOTHERAPY UNIT technical bid opening date being 02.07.2021 and as per decision taken in the purchase committee dated 03-09-2021 headed by M.D., RMSCL is under.

S. No.	Name of Bidder 's Firm	Name of item	Responsive or Non responsive	Reason for Non Responsive
1	Nice Neotech Medical Systems Pvt. Ltd.	Radiant Warmer, Radiant Warmer With Oxygen-Provision, Led Phototherapy Unit, T-piece Resuscitator	Responsive For Item Radiant Warmer , Radiant Warmer With Oxygen-Provision, Led Phototherapy Unit	Responsive
			Non responsive for Item T-piece Resuscitator	Technically not approved due to Model Name is not mentioned on product during demonstration..
2	Phoenix Medical Systems Private Limited	Radiant Warmer, Radiant Warmer With Oxygen-Provision, Led Phototherapy Unit,	Responsive For Item Radiant Warmer , Radiant Warmer With Oxygen-Provision, Led Phototherapy Unit	Responsive
3	AVI Healthcare Pvt. Ltd.	Radiant Warmer, Radiant Warmer With Oxygen-Provision, Led Phototherapy Unit, T-piece Resuscitator	Non responsive for Item Radiant Warmer, Radiant Warmer With Oxygen-Provision, Led Phototherapy Unit, T-piece Resuscitator	For Item Radiant Warmer Technically not approved during demonstration due to non availability of Pre-Heat mode and keyboard lock function
				For Item Radiant Warmer With Oxygen-Provision, Technically not approved during demonstration due to non availability of Pre-Heat mode and timer accuracy not optimal
				For Item Led Phototherapy Unit, Technically not approved during demonstration due to low irradiance
				For Item T-piece Resuscitator Documentary not approved due to non submission of European ce certificate from notified body/ USFDA certificate.
4	Medi Waves Inc.	Radiant Warmer, Led Phototherapy Unit	Non responsive for item Radiant Warmer, Led Phototherapy Unit	For Items Radiant Warmer, Led Phototherapy Unit Documentary Non-Responsive due to non submission of European ce certificate from notified body/ USFDA certificate
5	IRAYS MedTech Solutions Pvt. Ltd.	Radiant Warmer, Radiant Warmer With Oxygen-Provision, Led Phototherapy Unit	Non responsive Radiant Warmer, Radiant Warmer With Oxygen-Provision, Led Phototherapy Unit	For Items Radiant Warmer, Radiant Warmer With Oxygen-Provision, Led Phototherapy Documentary Non-Responsive because Firm participated as authorized dealer. Whereas, as per bid condition only Manufacturer/ Direct Importer is allowed to participate.

Handwritten signature and date: 17/9/21

6	FANEM MEDICAL DEVICES INDIA. PVT. LTD.	Radiant Warmer With Oxygen- Provision, Led Phototherapy Unit, T-piece Resuscitator	Responsive For item Led Phototherapy Unit, Non-responsive for Radiant Warmer With Oxygen- Provision, Non-responsive for T-piece Resuscitator	Responsive For item Radiant Warmer With Oxygen- Provision, Documentary Non-Responsive due to non-compliance of Past performance.
				For Item T-piece Resuscitator Documentary Non-Responsive due to non submission of European ce certificate from notified body/ USFDA certificate.
7	SOLOKRAFTS INDUSTRIES	Transcutaneous Bilirubinometer	Non-responsive for Transcutaneous Bilirubinometer	FOR Item Transcutaneous Bilirubinometer Documentary Non- Responsive due to non compliance of Past performance
8	GENWORKS HEALTH PVT LTD	Transcutaneous Bilirubinometer	Non-responsive for Transcutaneous Bilirubinometer	FOR Item Transcutaneous Bilirubinometer Documentary Non- Responsive due to non compliance of Past performance

Online financial bids of responsive bidders for above item will be opened on dated 21.09.2021 at 11:00 A.M.


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.
Jaipur