



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN: U24232RJ2011ISGC035067 E-Mail -edepmrmsc-rj@nic.in

No. F-8() RMSC/EPM/M-1/NIB-604/2021-22/ 1107

Dated: 27/09/21

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amendment in technical specification of bid document of Item Cardiac Monitor with Defibrillator, under No. F-8() RMSC/EPM/M-1/2021-22/NIB-604/949 date: 06.08.2021

Revised Technical specifications of Cardiac Monitor with Defibrillator-

1. Defibrillator should be Bi-Phasic.
2. Defibrillator should be ready for AED facility and non invasive pacing.
3. Should have a high resolution colour TFT/LED display of minimum 6 inch or more.
4. Should have direct trim knob and direct function keys for mute and freeze.
5. Should have energy levels for defibrillation up to 200 joules or more.
6. Should be mains and battery operated. Internal battery should provide backup operation up to 2 -3 hours in monitoring or at least 90 defibrillation shocks should be delivered from fully charged battery.
7. Integrated external re-usable adult and paediatric paddles for defibrillation.
8. Should have non synchronised and synchronised cardio version.
9. Facilities of ECG pickup from paddles in case of ECG electrodes are not connected to the Defibrillator.
10. The charging time should be less than 7 seconds for charging up to 200 joules.
11. Should have integrated Printer.
12. Quoted Model Should be European CE with notified body/USFDA approved.
13. Guarantee: Three years on equipment from the date of installation and 06 months on accessories.
14. CMC: CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
15. Equipment should be supplied with compatible 3/5 lead ECG cable (2 Nos.) & Operating / user Manual (1 No.)
16. Installation will be done by supplier free of cost.
17. Firm should have service engineer network in Rajasthan.
18. Demonstration of equipment is must for technical approval.
19. The firm should submit technical compliance sheet along with catalogue as per technical specifications mentioning the make & model of quoted item in the Technical bid.

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

27

Executive Director (EPM)
RMSCL, Jaipur