



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail: edepmrmsc-ri@nic.in

CIN:U24232RJ20115GC035067

Website: www.rmsc.health.rajasthan.gov.in

F-8() RMSC/EPM/M-4/NIB-813/ U25

Date: - 27/2/24

Clarification/Corrigendum/Addendum

Subject: Amendment in technical specification, item OAE ANALYZER under F-8() RMSC/EPM/M-4/NIB-813/2023-24/ 910 Dated 21.12.2023

REVISED TECHNICAL SPECIFICATIONS OF OAE ANALYZER

- 1. Power Supply: 220 Volts, 50Hz
- Type of OAE: TEOAEs, DPOAEs.
- Frequency range of TEOAE should be 4000 Hz or above and for DPOAE should be 3. 5000 Hz or above.
- Stimulai intensity level up to 80 dB SPL for TEOAEs and 70 db SPL or above for DPOAEs.
- Automatic F1 and F2 ratio calculation. 5.
- 6. Low battery warning
- 7. Serviceable probes; probe mount
- 8. Cleaning kit
- 9. Printer with minimum 10 paper roll.
- 10. Power supply: Rechargeable Battery operated
- 11. Can be used as stand alone or with a PC
- 12. Self-calibration
- 13. Separate eartips for infants and adults (50 no adults & 50 no. infant/pediatric).
- 14. Separate probe head/coupler 3 no's should be supplied free of cost with each machine.
- 15. Data Analysis should be fully automated (it should display cross spectrum, DP Gram, pass/Refer test report)
- 16. The equipment should be European CE (Notified body) /USFDA approved.
- 17. Should have safety certificate from a competent authority CE or valid details of electrical and functional safety test. Copy of the certificate/test report shall be produced along with the technical bid.
- 18. Three years Guaranty to be given with the equipment.
- 19. Minimum of two routine maintenance services/inspections to be provided annually.
- 20. CMC: CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
- 21. Price quoted are inclusive of their installation at site.
- 22. Final technical approval in equipment/Supply as per demonstration.

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Executive Director (EPM)

Signature wal

Digitally signed by Dr. kash Alha Designation Executive Director Date: 2024.02.27 17.07:56 IST

Reason: Approve

