



## Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur  
Phone No.-2223887, Fax no.-0141-2228065

CIN:U24232RJ2011SGC035067

E-Mail : edepmrmc-rj@nic.in

Website: www.rmhc.health.rajasthan.gov.in

F-8(279) RMSC/EPM/M-2/NIB-660/2021-22/1718

Date:- 24/02/2022

### Clarification/Corrigendum/Addendum

**Subject:- No. F-8(279) RMSC/EPM/M-2/NIB-660/2021-22/1579 Dated 07.01.2022 for item ICU Bed**

In Reference to subject cited above and NIB-660 the various representations received from the firms and issues raised by the Bidders are under consideration by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

#### **1. Clarification Technical Specification of ICU Bed**

S.No.	Published as below	Read as below
17.	Caster Wheels: The bed shall be provided with <b>Diagonally</b> locking 125mm diameter wheels and directional locking of wheels shall have thread guards; the castor body shall be made of non-rusting material. Corner Buffers: Four nos. rubber buffer shall be fitted to the four corners.	Caster Wheels: The bed shall be provided with <b>Diagonally</b> locking 125mm diameter wheels and <b>Diagonally</b> locking of wheels shall have thread guards; the castor body shall be made of non-rusting material. Corner Buffers: Four nos. rubber buffer shall be fitted to the four corners.

**Note:-** Please note that all above amendments/corrigendum is technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Executive Director (EPM)  
RMSCL, Jaipur