

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN: U24232RJ2011SGC035067 E-Mail - edepmrmsc-ri@nic.in

No. F-8() RMSC/EPM/M-1/NIB-645/2021-22/ 1796

Dated: 29/8/2022

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amendment in technical specification of bid document of item Dialysis Machine (Haemodialysis Machine), under No. F-8() RMSC/EPM/M-1/NIB-645/2021-22/1672 date: 08.02.2022

In reference to subject cited above NIB-659, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

Revised Technical Specification for Dialysis Machine (Haemodialysis Machine):-

1. Blood pumps to achieve a wide range of flow adaptable to standard A-V blood lines.
2. Heparin pump with adaptability of various sizes of syringes up to 20ml with pump flow rate from 1-10ml min and bolus infusion up to 5ml/min.
3. It should have sequential ultra filtration facility. It should have arterial, venous & trans-membrane pressure monitoring facility.
4. Functional air bubble detector.
5. It should have dialysate flow rate range from 300 to 800ml/min
6. Variable conductivity setting between 12.5 to 16ms/cm
7. Functional blood leak detector.
8. Dialysate temperature regulator that has a range of temperature 35°C to 39°C
9. Safety device: Functioning Alarms, Venous Blood Clamp.
10. It should have control & monitoring screen with minimum 10" coloured TFT/LCD Display (Screen). Display should have screen with high clarity & easily readable.
11. It should have hot rinsing and hot chemical disinfection facility (Temp. up to 80°C) with recirculation system.
12. It should have in-built battery backup of minimum 30 minutes for complete extracorporeal blood system during power failure.
13. Non-Invasive patient blood pressure monitoring.

General Terms & Conditions:


1. The product should be USFDA/European CE (Notified Body) certified. (Valid documentation should be submitted in technical bid).
2. Demonstration of the machine is must for final technical approval.
3. List of consumables, accessories & spare parts as per requirement of the equipment should be provided by supplier in the technical bid.
4. The company have excellent service backup with resident engineer based in major cities of Rajasthan,
5. Installation and training for medical/technical staff to getting acquaint with machine at consignee will be done by the supplier free of cost. Operating manual/training material (1set) will be supplied by the supplier free of cost at the time of installation.

g.



6. 3 year warrantee on complete product from the date of installation.
7. CMC: CMC will be given @ 5% exclusive GST (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of warrantee period of three years.
8. Four Preventive Maintenance services during warrantee and CMC period are essential. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted product along with catalogue in the Technical bid.

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur



