



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmse-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: <http://rmse.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-3/20-21/NIB-547/189

Dated: 06/11/2020

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Amended technical Specifications and revised bid submission date of the rate contract for Oxygen Concentrator for under NIB No. F-8() RMSC/EPM/M-3/20-21/NIB-547/169 dated 28.10.2020

Revised Technical Specifications of Oxygen Concentrator

1. Oxygen concentrator to provide oxygen from ambient air.
2. Oxygen flow measured at the flow meter.
3. Sound level should be ≤ 55 dB
4. Maintenance free rotary proppet valve.
5. Oxygen purity should be $88\% \pm 2\%$ or more.
6. **Should have dual Oxygen output/ single output with splitter, having individual oxygen flowmeter.**
7. **The overall oxygen flow should have at least 5 LPM or more.**
8. **Pressure range should be 5 PSI or more.**
9. **Oxygen nasal cannula** of 2m length (02no.) with autoclavable polycarbonate humidifier bottle and two cabinet filters should be provided with each oxygen concentrator.
10. Supplied with:-
 - i. User manual with trouble shooting guidance, in English.
 - ii. Technical manual with maintenance and first line technical intervention instructions, in English.
11. **Power requirements: 220-240 V/50 Hz.**
12. Manufacturer of the device should have ISO 9001:2015 and ISO 13485:2016 certified manufacturer. Related documents/certification shall be submitted along with technical bid.
13. Oxygen concentrator device should be certified ISO 80601-2-69:2014 certified.
14. Oxygen failure or Low Oxygen Alarm facility should be available for below 80% oxygen.
15. Device should have USFDA approval / CE certificate issued by a notified body registered in the European commission, under relevant category. Related documents / certification shall be submitted along with technical bid.
16. **Guarantee: Three years on equipment from the date of installation.**
17. CMC: CMC may be given @ 5 % exclusive GST (as applicable) and yearly escalation of 5 % previous year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of two years.
18. The service engineer should be based in Rajasthan.
19. Demonstration of equipment is must for technical approval.
20. Installation will be done by supplier free of cost.
21. The company should mention the make & model of the quoted goods and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications






Revised bid schedule:-

E-bids are invited as per following revised time schedule:-

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
<u>09.11.2020</u> <u>11:00 a.m.</u>	<u>09.11.2020</u> <u>6:00 p.m.</u>	<u>10.11.2020</u> <u>03:00 p.m.</u>	<u>12.11.2020</u> <u>11:00 a.m.</u>	<u>12.11.2020</u> <u>6:00 p.m.</u>	<u>13.11.2020</u> <u>03:00 p.m.</u>

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur