

**Rajasthan Medical Services Corporation Limited, Jaipur**

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmse-rj@nic.in

CIN:U24232RJ2011SGC035067Website: http://rmse.health.rajasthan.gov.in

No. F-8( ) RMSC/EPM/M-3/20-21/NIB-514/ 45

Dated: 23.09.2020

**CLARIFICATION/CORRIGENDUM/ADDENDUM**

**Sub:-Amended Technical Specification and bid submission date of the rate contract for Oxygen Concentrator for under NIB No. F-8( ) RMSC/EPM/M-3/20-21/NIB-514/25 dated 16.09.2020**

**Revised Technical Specifications of Oxygen Concentrator**

1. Oxygen concentrator to provide oxygen from ambient air.
2. Oxygen flow measured at the flow meter.
3. Sound level should be  $\leq 55$ dB
4. Maintenance free rotary proppet v
- 5.
6. alve.
7. Oxygen purity should be  $88\% \pm 2\%$  or more.
8. Oxygen output should be at least 5 LPM or more.
9. Pressure range should be 6 - 8 PSI.
10. Oxygen tube of 2m length (01no.) with autoclavable polycarbonate humidifier bottle and two cabinet filters should be provided with each oxygen concentrator.
11. Supplied with:-
  - i. 1 x spare set of tubing.
  - ii. 1 spare sets of humidifier bottle.
  - iii. Two spare sets of cabinet filters.
  - iv. User manual with trouble shooting guidance, in English.
  - v. Technical manual with maintenance and first line technical intervention instructions, in English.
12. Power requirements: 220 V/50 Hz.
13. Device is produced by ISO 9001 and ISO 13485 certified manufacturer. Related documents/certification shall be submitted along with technical bid.
14. Device should be certified ISO 80601-2-69:2014.
15. Oxygen failure or Low Oxygen Alarm facility should be available for below 80% oxygen.
16. Device should be safety certified according CE / USFDA. Related documents/certification shall be submitted along with technical bid.
17. Guarantee: Two years on equipment from the date of installation.
18. CMC: CMC may be given @ 5 % exclusive GST (as applicable) and yearly escalation of 5 % previous year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of two years.
19. The service engineer should be based in Rajasthan.
20. Demonstration of equipment is must for technical approval.
21. Installation will be done by supplier free of cost.
22. The company should mention the make & model of the quoted goods and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications



**Revised bid schedule:-**

E-bids are invited as per following revised time schedule:-

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
<u>24.09.2020</u>	<u>24.09.2020</u>	<u>25.09.2020</u>	<u>29.09.2020</u>	<u>29.09.2020</u>	<u>30.09.2020</u>
<u>11:00 a.m.</u>	<u>6:00 p.m.</u>	<u>03:00 p.m.</u>	<u>11:00 a.m.</u>	<u>6:00 p.m.</u>	<u>03:00 p.m.</u>

**Note:-**

1. Last Date and Time for physical submission of DD/BC/BG/Challan for Bid Document Fee, RISL Fee, Bid Security Money, Original Affidavits/Certificates: 30.09.2020 up to 1.00 PM
2. Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

23/9/20  
Executive Director (EPM)  
RMSCL, Jaipur