



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No. 2223887, Fax no.- 0141 2228065

E-Mail: edepmrmsc-rj@nic.in

CIN: U24233RJ2011SGC035067

Website: www.rmsh.health.rajasthan.gov.in

GEM/2026/B/7172077 / 507

Date:- 10/07/2026

Clarification/Corrigendum/Addendum

Subject: Amendment in technical specification, item Ceiling OT Light (under GEM/2026/B/7172077 Dated 30-01-2026)

Revised Technical Specification of Ceiling OT Light

1. Surgical light unit with Electronics control, for homogeneous and shadow less light with minimum temperature rise.
2. The highly counter balanced mechanism to fix the dome at any position, easy maneuverable free up & down & swing movements & freely rotatable at any angle (360 degree).
3. Light intensity should be adjustable from control panel available on dome.
4. Focusing should be adjustable from control panel or manually by sterile handle. Sterile handles two numbers extra should be supplied with light.
5. Light intensity at a distance of 1 meter for each dome at least 1,60,000 lux and Light field diameter shall be above 20 cm or better.
6. Life of LEDs minimum 40,000 working hrs or more.
7. Color temperature Variable (3800 to 1500 degree K)
8. Light dome shall be compatible for laminar air flow.
9. Power supply: 230 Volts AC, 50Hz.
10. Product should be CE certified & manufacturer should be ISO 13485 certified. (Valid documentation should be submitted in technical bid).
11. **Equipment must have safety requirement of IEC 60601-1 and IEC 60601-2-41. (Documentation should be submitted in technical bid).**
12. Installation will be done free of cost by supplier with all Civil, Mechanical and Electrical works required.
13. Guarantee: Three years from the date of installation.
14. CMC: CMC will be given @ 5% plus GST (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of three years, if required.
15. The service provider should be based in Rajasthan.
16. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
17. Final technical approval after demonstration

Note:- All other terms & conditions remains the same.

RajKaj Ref No.:
23329351

Director (P)
RMSCIL, Jaipur



Digitally signed by
ANJU OMPRAKASH
Designation: Director
Date: 2026-07-10
12:10:43 IST