



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN: U24232RJ2011SGC035067 E-Mail - gdenmrmsc-ri@nic.in

No. F-8() RMSC/EPM/M-1/NIB-809/2023-24/ 699

Dated: 21/11/23

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amendment in technical specification and bid schedule of bid document of item Air Purifier System, under No. F-8() RMSC/EPM/M-1/NIB-809/2023-24/579 date: 29.09.2023

A. Revised Technical Specification Air Purifier System:-

1. It shall have 7 stage air purification technologies that include:
 - 1) Pre-Filter washable
 - 2) Anti Bacterial (Easily Replaceable)
 - 3) HEPA Filter (Easily Replaceable)
 - 4) Activated Carbon Filter (Easily Replaceable)
 - 5) Tio2 Photo Catalyst Filter (Easily Replaceable)
 - 6) Ultra Violet Lamp-2 Nos.
 - 7) Ionizer
2. Airflow : 225-250 m3/hr
3. HEPA Filter/ULPA Filter: 0.3 micron @97%, 1 micron @99.97%
4. Voltage : 220 V
5. Power Consumption : 40-45 Watt
6. Min Noise Level : 35 db (± 5db)
7. Max Noise Level : 45 db (± 5db)
8. Timers : 1 to 12 Hours & Manual or IOT
9. Weight : 4.5-5.5 kg
10. Remote : Yes
11. Size : 330x180x500mm (±50 mm for L,W & H)
12. Filters need to be easily replaceable
13. COMPANY SHOULD BE ISO 9001:2015
14. Guarantee 1 Year

B. Revised Bid Schdeule—

E-bids are invited as per following revised time schedule:-

Existing Dates				Extended Dates			
Last date & time for sale of bid form	Last date & time of receipt of bid form	Last date of submission of bid document fees	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Last date of submission of bid document fees	Date & time of opening of technical bid
<u>02.11.2023</u> <u>11:00 a.m.</u>	<u>02.11.2023</u> <u>6:00 p.m.</u>	<u>03.11.2023</u> <u>02:00 p.m.</u>	<u>03.11.2023</u> <u>03:00 p.m.</u>	<u>16.11.2023</u> <u>11:00 a.m.</u>	<u>16.11.2023</u> <u>6:00 p.m.</u>	<u>17.11.2023</u> <u>02:00 p.m.</u>	<u>17.11.2023</u> <u>03:00 p.m.</u>



Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur 