



**RAJASTHAN MEDICAL SERVICES CORPORATION LTD.**  
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail -[edepmrmisc-ri@nic.in](mailto:edepmrmisc-ri@nic.in)

No. F-8( ) RMSC/EPM/M-3/2021-22/NIB-597/ 704

Dated: 10/05/2021

**CLARIFICATION/CORRIGENDUM/ADDENDUM**

**Subject: Amended technical specifications and revised bid schedule of bid document of item 1. Oxygen Concentrator (5 LPM) 2. Oxygen Concentrator (10 LPM) under No. F-8( ) RMSC/EPM/M-3/NIB-597/2021-22/ 657 date: 01.05.2021**

In reference to subject cited above NIB-597, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

| OXYGEN CONCENTRATOR 5LPM  |  |  |
|---------------------------|--|--|
| S.No                      | Existing Technical Specification   | Amended Technical Specification  |
| 1.                        | Should have oxygen purity Indicator and alarm if oxygen purity level decreases below 85%.        | Should have oxygen purity Indicator and alarm if oxygen purity level decreases below 82%.          |
| OXYGEN CONCENTRATOR 10LPM |  |  |
| 1.                        | Should have oxygen purity Indicator and alarm if oxygen purity level decreases below 85%.        | Should have oxygen purity Indicator and alarm if oxygen purity level decreases below 82%.          |
| 2.                        | Should have 02 (two) flowmeters having flow from 0-5 litres per minute with variation of 0.5LPM. | Should have 02 (two) flow meters having flow from 0-10 liters per minute with variation of 0.5LPM. |

**Revised bid schedule:**

E-bids are invited as per following revised time schedule:-

| Existing Dates                 |                                  |                                  | Extended Dates                 |                                  |                                  |
|--------------------------------|----------------------------------|----------------------------------|--------------------------------|----------------------------------|----------------------------------|
| Last Date for Sale of Bid Form | Last Date of Receipt of Bid Form | Date of Opening of Technical Bid | Last Date for Sale of Bid Form | Last Date of Receipt of Bid Form | Date of Opening of Technical Bid |
| 1                              | 2                                | 3                                | 4                              | 5                                | 6                                |
| <u>10.05.2021</u>              | <u>10.05.2021</u>                | <u>11.05.2021</u>                | <u>12.05.2021</u>              | <u>12.05.2021</u>                | <u>13.05.2021</u>                |
| <u>11:00 a.m.</u>              | <u>6:00 p.m.</u>                 | <u>03:00 p.m.</u>                | <u>11:00 a.m.</u>              | <u>6:00 p.m.</u>                 | <u>03:00 p.m.</u>                |

**Note:-** 1. It is also clarified that information of award of contract shall be communicated to all participating bidders on the website [www.rmhc.health.rajasthan.gov.in](http://www.rmhc.health.rajasthan.gov.in) and [sppp.raj.nic.in](http://sppp.raj.nic.in). Please note that individual bidder will not be intimated.

2. Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Executive Director (EPM)  
RMSCL, Jaipur