



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail -edepmrmsc-ri@nic.in

No. F-8()RMSC/EPM/M-6/NIB-542/ 2020-21/ 817

Dated: 24/06/2021

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amended Specification, Bid schedule of item 1:Laparoscope for single puncture laparoscopic sterilization with CO2 insufflators and cold light source 2. Laparoscope set for training with CO2 insufflators under No. F-8)RMSC/EPM/M-6/NIB-542/ 2020-21/237 date: 01.12.2020

In reference to subject cited above NIB-542, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

A-Amended in technical specification of item Laparoscope for single puncture laparoscopic sterilization with CO2 insufflators and cold light source as below:-

Single Puncture Laparoscope for Laparoscopic Sterilization with CO2 Insufflators and Cold Light Source		
S.N.	Tender Specification	Amended Specification
1.	Telescope: Telescope 30 degree with parallel/straight eye piece, 10 mm diameter. Should have 6 mm instrument channel built-in ring applicator Working length of 270-275 mm/10mm 0 degree telescope. Fibre optic light transmission incorporated, should be compatible with the commonly available light cables (necessary adaptors should be provided) Can be sterilized by autoclaving, gluteraldihyde solutions and formalin chamber.	Telescope: Telescope 0 degree with parallel/straight eye piece, 10 mm diameter. Should have 6 mm instrument channel built-in ring applicator Working length of 270-275 mm/10mm 0 degree telescope. Fibre optic light transmission incorporated, should be compatible with the commonly available light cables (necessary adaptors should be provided to Provide compatibility to light cables of all make and to introduce 5mm hand instruments through the adapter in case of need of surgical interventions of the fallopian tubes) Can be sterilized by autoclaving, gluteraldihyde solutions and formalin chamber.
2.	Compatibility & Core items: The core laparoscopy items such as Telescope, Trocars& Cannula, Ring Applicator, Veress Needle, CO ₂ insufflators, Light Source and Fiber Optic light cable etc. should be from single manufacturer for system compatibility. Item like formalin chamber, cleaning kit, briefcase, tray for sterlization, ring applicator , UPS, LED monitor, etc may be from different manufacturer but should be compatible with core item.	Compatibility & Core items: The core laparoscopy items such as Telescope, Trocars& Cannula, Ring Applicator, Veress Needle, CO ₂ insufflators, Light Source and Fiber Optic light cable etc. should be from Same manufacturer for system compatibility. Item like formalin chamber, cleaning kit, briefcase, tray for sterlization, UPS, LED monitor, etc may be from different manufacturer but should be compatible with core item.

[Handwritten signature]
24/6/21



B-Revised bid schedule:-

- E-bids are invited as per following revised time schedule:-

Existing Dates			Extended Dates		
Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid
28.06.2021 06.00 p.m.	28.06.2021 06.00 p.m.	29.06.2021 03.00 p.m.	05.07.2021 06.00 p.m.	05.07.2021 06.00 p.m.	06.07.2021 03.00 p.m.

Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Executive Director (EPM)
RMSCL, Jaipur