



## Rajasthan Medical Services Corporation Ltd.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN:U24232RJ2011SGC035067 E-Mail -rmisc@nic.in, edepmrmisc-ri@nic.in

No. F-8() RMSC/EPM/M-4/NIB-510/2020-21/ 834

Date: 25/06/2021


### Office Order

List of bidders declared responsive/non responsive as per documentary & technical evaluation in bid number F-8()RMSC/EPM/M-4//NIB-527/2020-21/114 Dated: 14.10.2020 for item Electrical Suction Machine 2. Syringe Needle Destroyer 3. Weighing Machine(Adult) 4. Weighing Machine (Neonatal) 5. Suction Machine (Foot Operated) technical bid opening date being 19.01.2021 and as per decision taken in the purchase committee dated 28.05.2021 headed by M.D., RMSCL is under.

| S. No. | Name of Bidder 's Firm                                       | Name of quoted item  | Responsive / Non responsive  | Reason for Non Responsive   |
|--------|--|--|--|---|
| 1      | M/s Medicare Products. Inc. New Delhi-110015(India)          | Syringe Needle Destroyer   | Responsive   | Responsive  |
| 2.     | M/s Ronak Scale Co. Manufacture Electronic Scale Savarkundla | Weighing Machine (Neonatal)  | Non-Responsive   | तकनीकी परीक्षण (Demo) के दौरान फर्म द्वारा प्रस्तुत आईटम में auto off feature and stamping from legal metrological deparment on weighing machine (Neonatal) नहीं होने के कारण फर्म Demo में Not approved रही ।  |
| 3.     | SRK Modular Furniture Co. M.I. Road Jaipur(Raj)              | 1. Syringe Needle Destroyer<br>2. Weighing Machine(Adult)<br>3. Weighing Machine (Neonatal)        | Responsive for Syringe Needle Destroyer<br><br>Non Responsive for Weighing Machine(Adult) and Weighing Machine (Neonatal)        | 1. तकनीकी परीक्षण (Demo) के दौरान फर्म द्वारा प्रस्तुत आईटम Syringe Needle Destroyer तकनीकी समिति ने as per Specification पाया गया।<br>2. फर्म द्वारा आईटम Weighing Machine(Adult) का डेमोनस्ट्रेशन नहीं दिया गया।<br>3. तकनीकी परीक्षण (Demo) के दौरान फर्म द्वारा प्रस्तुत आईटम Weighing Machine (Neonatal) में Non Removable Tray and Mattress नहीं पाया गया । प्रस्तुत आईटम में कोई Make/Model नहीं पाया गया। अतः फर्म को Syringe Needle Destroyer के लिए Approved और Weighing Machine(Adult) एवं Weighing Machine (Neonatal) के लिए Non Approved किया गया। |
| 4.     | M/s Hardik Medi Haryana                                      | Weighing Machine(Adult)  | Responsive   | Responsive  |
| 5      | M/s United Surgical Industries, Delhi                        | 1. Syringe Needle Destroyer<br>2. Electrical Suction Machine<br>3. Suction Machine (Foot Operated) | Responsive for Syringe Needle Destroyer and Suction Machine (Foot Operated)<br><br>Non Responsive for Electrical Suction Machine | 1. तकनीकी परीक्षण (Demo) के दौरान फर्म द्वारा प्रस्तुत आईटम Syringe Needle Destroyer एवं Suction Machine Foot Operated तकनीकी समिति ने as per Specification पाया गया।<br>2. तकनीकी परीक्षण डेमो के दौरान फर्म द्वारा प्रस्तुत आईटम Electrical Suction Machine में SMPS नहीं होने के कारण डेमो में Non Approved की गयी   |

|   |  |                                |            |            |
|---|--|--------------------------------|------------|------------|
| 6 | M/s Medi Waves Inc.<br>Sonapat,<br>Haryana | Weighing Machine<br>(Neonatal) | Responsive | Responsive |
|---|--|--------------------------------|------------|------------|

Online financial bids of responsive bidders for above item will be opened on dated 29.06.2021 at 11:00 A.M.

  
**Executive Director (EPM)**  
**Rajasthan Medical Services Corporation Ltd.**  
**Jaipur**