

## Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmsc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: <http://rmsc.health.rajasthan.gov.in>

No. F-8( ) RMSC/EPM/M-3/21-22/NIB-602/855

Dated: 29/06/2021

### CLARIFICATION/CORRIGENDUM/ADDENDUM

**Sub:- Amended technical specifications of the rate contract for Liquid Medical Oxygen (LMO) for under NIB No. F-8( ) RMSC/EPM/M-3/21-22/NIB-602/767 dated 07.06.2021.**

In Reference to subject cited above and NIB-602, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued:-

### Revised Technical Specification of Liquid Medical Oxygen [LMO]

Existing Specification	Amended Specification
3. Measured oxygen contains should not be less than 99.6% v/v	Measured oxygen contains should not be less than 99% v/v
6.Should not measure more than 30 ppm v/v carbon dioxide.	Should not measure more than 300ppm v/v carbon dioxide.
12. Manufacturing plants of LMO should be ASTM certified and ISO 9001 and 14001 compliant.	Manufacturing plants of LMO should be ASTM certified/ISO 9001/14001 compliant.
15. Random sampling will be done at the institute and pilferage or less quantity will be recovered at panel rate i.e. double the supply rate. If percentage of oxygen is less than 99.6% then price will be reduced accordingly	Random sampling will be done at the institute and pilferage or less quantity will be recovered at panel rate i.e. double the supply rate. If percentage of oxygen is less than 99% then price will be reduced accordingly

**In the tender form, Section-VII SPECIAL CONDITIONS OF RATE CONTRACT(SCC) Clause no. 5 is deleted.**

**Note:-** Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.

  
Executive Director (EPM)  
RMSCL, Jaipur