

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228068 CIN: U24232RJ20118GC035067 E-Mail edepmruise-rjænic.in
No. F-8() RMSC/EPM/M-1/NIB-720/2022-23/ 8 83

Dated: 16 1 23

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amendment in Technical Specification, GENERAL CONDITIONS RATE CONTRACT (GCC), BID DATA SHEET (BDS), Revised bid schedule of bid document of item 16 Slices CT Scan Machine, under No. F-8() RMSC/EPM/M-

1/NIB-720/2022-23/637 date: 22.11.2022

1-Revised Technical Specifications:

Technical Specification	Technical Specification of 16 Slice CT Scan						
Parameter/ Feature	THRESHOLD LIMIT/ Range						
Gantry							
Scan time for 360 Degree rotation	0.75 Seconds or Less						
Minimum Slice Thickness	0.7 mm or Less 5.50 or more.						
Fov in acquisition Mode(in cm)							
Aperture	70 cm or more						
X-Ray Section							
System X-ray Power	32 KW or more						
mA range	(13 to 240mA) or more.						
Heat Storage capacity in MHU	3.5 or more.						
X-ray tube cooling rate in KHU per min	700 or more						
Detector							
No of Detector Rows	16 or more.						
Defective element/channels per row	650 or more.						
Patient Couch							
Road bearing capacity	160 Kg or more						
Range of metal free scan	125 cms or more.						
Spiral/Helical Section							
spiral Capability	100 Second or more.						
Length of continuous spiral scan	100 cms or more.						
Computer System							
Ram size in GB	32 GB or more						
No. of the monitor	1						
Monitor size in inches	19 or more.						
Display Matrix	1024x1024 or more.						
Mention Hard disk capacity for image and raw date	v 01 TB or more.						
Image storage capacity	100000 images or more.						
Advanced Work station Ram	32GB or more.						
Resolution							
High contrast resolution in lp/cm	15 or more.						
Low contrast resolution in mm at 0.3%	Low contrast resolution in HU at 20 Phantomm CATPHAN:3 HU or more						
DRY Camera							
Laser Printer with 3 minimum ports	16 bits and 500dpi						
Throughput per hour	70 or more.						
Must have specification							
Parameter/ Feature	Yes/No						

Letin



N	The last of the same property and some some some some some some some some	1
	Gantry Normal Sit/Digital Tilt/Flex e-Tilt	-
	Remote till/Ingital tild/fex c 1	-
	With user control panels on either sides	
	3 D positioning laser lights	
	Dual focus X-Ray Tube	
	X-ray tube cooling unit in built in gantry	-
	Detectors	
	Soli State	-
	Patient Couch	-
	Remote UP/Down, FWD/BWD option	
	Spiral/Helical Sanction	
	Ability to track contrast medium to trigger scan	
	Dose reduction scan	-
	Computer System	-
	Multitasking processors and menu driven	
	platform	
1	OEM Additional work station with all the	
	software available in console	-
	Archiving facility of DVD & CD Main console	
	Capability from HIS-BIS interface	
	PACS interface without any new hardware or	
	software	
	DICOM3.0	
	PC based connectivity	
	OEM Advanced work station	
	Image Processing	
	3D volume rendering	
	MIP	
	CT angiography	
	Artifact suppressions	
	Pre set 3D reconstruction & display protocols	
	Auto Bone removals	
	CT based DSA attachment	
11	CT Perfusion	
	Dental Scan	
	Bone removal	
-	MPR	
1	ROI	
1	Volume calculation	
	CT Number Measurement	
-	Window Width	
	Window Level Tengaryan Dieplay	
	Topogram Display CINE Display	
	CINE Display	
1	HRCT Lung Departie Sam	
1	Dynamic Scan Direct disclay of MPR images Immediately offer	
	Direct display of MPR images Immediately after Scan	
1	Accessories	
	Single head contrast injector (300 PSI) of reputed make	
	Others	
1		
1	facility Model launched in 2014 or after	
1		-
1	Standard and Quality Certification EUROPEAN CE/US FDA	
1	TOROLLAN CEJUS FUA	

Scanned with Camsca



Accessories to be included for 16 Slice CT Scanner

Lead Glass of size-100 X 150 cm-1 no Lead Apron Hanger -2 nos. Zero lead Apron -6 nos.

- i. Online UPS from reputed manufacturer having at least 30 minute back up with appropriate KVA of the entire system. The backup should be such that scan can be done uninterrupted
- ii. All patient positioning accessories including head rest.
- iii. Anaesthesia Workstation with Ventilator having USFDA/CE Certified product- 1 nos.
- iv. Single head contrast injector (300 psi) Syringe- 100 nos.

Later

2-SECTION-V: BID DATA SHEET (BDS)

Existing	Amended		
Point No. 22: This bid do not have provision for Parallel Rate Contract (PRC).	Point No. 22: This bid has provision for parallel rate contract in the ratio as decided by MD RMSCL, if required.		

3-SECTION II:-GENERAL CONDITIONS RATE CONTRACT (GCC)

Addition in 45-Performance Security (PS) and agreement:

75A. Additional Performance Security.- (1) In addition to Performance Security as 75. rule in specified Performance Security shall also be taken from the successful bidder in case of Performance Additional bid. The unbalanced Security shall be equal to fifty percent of Unbalanced Bid Amount. The Additional Performance Security shall be deposited in lump sum by the successful bidder before execution of Agreement. The Additional Performance Security shall be deposited through e-Grass, Demand Daft, Banker's Cheque, Government Securities or Bank Guarantee.

Explanation: For the purpose of this rule,-

- i. Unbalanced Bid means any bid below more than fifteen percent of Estimated Bid
- ii. Estimated Bid Value means value of subject matter of procurement mention in bidding documents by the Procuring Entity.
- iii. Unbalanced Bid Amount means positive difference of eighty five percent of Estimated Bid Value minus Bid Amount Quoted by the bidder.
 - (2) The Additional Performance Security shall be refunded to the contractor after satisfactory completion of the entire work. The Additional Performance Security shall be forfeited by the Procuring Entity when work is not completed within stipulated period by the contractor. Provision for 'Unbalanced Bid' and 'Additional Performance Security' shall be mentioned in the Bidding Documents by the Procuring Entity.

4-Revised Bid Schdeule--

E-bids are invited as per following revised time schedule:-

E-D	ias are invi	ted as per 10	mowing revis	cu time sen	caure.		
Existing Dates			Extended Dates				
Last date & time for sale of	Last date	Last date	Date & time of opening of technical bid	& time for	Last date & time of receipt of bid form		Date & time of opening of technical bid





मुख्यमंत्री निःशुल्क जाँच योजना

_								
			document				document	
			fees				fees	
	17,01,2023	17.01.2023	18.01.2023	18.01.2023	31.01.2023	31.01,2023	01.02.2023	01.02.2023
	11:00 a.m.	6:00 p.m.	02:00 p.m.	03:00 p.m.	11:00 a.m.	6:00 p.m.	02:00 p.m.	03:00 p.m.

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Executive Director (EPM)
RMSCL, Jaipur

Letin