



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur
Phone No.-2223887, Fax no.-0141-2228065 E-Mail : edepmrmisc-rj@nic.in
CIN:U24232RJ2011SGC035067 Website: www.rmhc.health.rajasthan.gov.in
F-8()RMSC/EPM/M-2/NIB-633/2021-22/ 9245 Date:- 2.12.21

Clarification/Corrigendum/Addendum

Subject:- Revised Technical Specification for NIB No. F-8()RMSC/EPM/M-6/NIB-633/2021-22/1138 Dated: 30.09.2021 for item Multi Para Monitor.

In Reference to subject cited above and NIB-633, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

Revised Technical Specifications of Multi Para Monitor

- 1) It should having ECG, SpO2, NIBP, Respiration and two Temperature as standard parameters.
- 2) Should have facility to display 7 lead ECG, RR, HR, SpO2, NIBP, and dual Temperature.
- 3) Should display 6 or more waveforms of selected parameters simultaneously, should have colour coding for different waveforms.
- 4) Should have inbuilt battery backup of at least three hours.
- 5) Display:Color LED/LCD touch screen display size of 12" or more.
- 6) Monitor should have both touch screen and knob to operate.
- 7) Should have dual temperature monitoring either in 0Celsius or 0Fahrenheit.
- 8) Should have facility for displaying multi - screen configurations.
- 9) Should be able to store & display at least 96 hours of graphical trends of all parameters
- 10) Should be suitable for monitoring adult & pediatric & neonate patients.
- 11) The SpO2 technology should be Masimo/Nellcor/TruSignal and it should also sense in hypotension, shivering & motion.
- 12) Should have oscillometric technology for measurement of NIBP with Auto, START and Manual modes.
- 13) Should have different patient type selection.
- 14) The pulse rate should be displayed either with ECG or SpO2.
- 15) The respiration rate should be calculated through Impedance method.
- 16) Should be able to analyze arrhythmias & ST segment changes.
- 17) Memory should not wipe off when the power is turned off.
- 18) Should be able to give visual & audible alarms with three levels of volume adjustment.
- 19) Should have facility of connecting to the central monitoring station having facility for real time monitoring through Ethernet cable or Wireless Connectivity.

20) Monitor should have VGA port for External Display

21) Wall mount stand should be provided free of the cost with the monitor.


22) Quoted model should be European CE (Notified Body)/USFDA approved.

23) Warrantee-Three years on equipment from the date of installation.

24) CMC: CMC will be given @ 5 % of Net Rate exclusive of GST (As applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Warrantee period of three years.

25) Equipment should be supplied with compatible 5 lead ECG cable (1 Nos.), SpO2 (Adult, Pediatric, Neonate) probe with extension cable (Each 1No.), NIBP Cuff with extension cable of Large Adult, Adult & Pediatric (2 each), Temperature probe (2 Nos.) & Operating / user Manual (1 No.).

All other terms & conditions remains the same.



Executive Director (EPM)
RMSCL, Jaipur