

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur
Phone No.-2223887, Fax no.-0141-2228065
CIN:U24232RJ2011SGC035067
No. F-8(338) RMSC/EPM/M-2/NIB-812/2023-24/

E-Mail : edepmrmse-rj@nic.in
Website www.rmhc.health.rajasthan.gov.in

Date:-

Clarification/Corrigendum/Addendum

Subject:- NIB No. F-8(338) RMSC/EPM/M-2/NIB-812/2023-24/908 Dated 20.12.2023 for item BP APPARATUS PROFESSIONAL

In Reference to subject cited above and NIB-812 The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

1. Revised Specification:-


AMENDMENT/REVISED TECHNICAL SPECIFICATIONS OF "OF BP APPARATUS PROFESSIONAL

1. Model(s) to be designed for professional use in hospital settings (Personal home-based models will not be considered).
2. Quoted model meets standards of AAMI/ANSI/ISO 81060-2:2013/BHS/ESH-IP(Valid document/certificate must be submitted in technical bid).
3. Pressure measurement range should be 60 to 290 mm Hg systolic, and 40 to 200 mm Hg diastolic.
4. Pressure measurement accuracy of ± 5 mm Hg.
5. Pulse rate to be displayed.
6. Should have LCD/TFT/LED Display of 69*64 mm or more.
7. Measurement Method: Oscillometric Measurement.
8. Standard Cuff Size: Adult Normal (2 Nos), Adult Large(1 Nos)& Pediatric Small (1 Nos) to be supplied free of cost with each machine.
9. Device should have internally rechargeable battery of More than 150 measurements.
10. Should have AC Adaptor.
11. Detachable cuff/sleeve.
12. Built-in surge protection to prevent damage to instrument in case of power surge.
13. Three years warranty including all spares, accessories, consumables and re-calibration (Bidder has to replace the unit if not repairable within warranty period).
14. Devices should include a temperature-stabilizing system, which allows for use in weather conditions (0°C to 50 °C).
15. Power Saving: Automatic power off if system idle for 5 minutes.
16. Minimum- Three years warranty including all spares, accessories, consumables and breakdown complaint must be provided within 24 Hours.

Signature valid



Digitally signed by Dr. Akash Alha
Designation: Executive Director
Date: 2024.03.07 10:14:55 IST
Reason: Approved



17. Quoted model should have US FDA/European CE with 4 digit notified body/BIS (Valid certificate must be submitted in technical bid).
18. Device should have Bluetooth connectivity.
19. Device should have cuff holder inbuilt or attached with device.
20. Device should have movement detection, Loose cuff detection, Average of last 3 readings, 7 Day AM/PM Averaging .
21. Copy of CDSCO License (if applicable) must be submitted in technical bid.
22. Low Battery indicator and error indicators should be available.
23. Carrying case/bag to be provided free of cost with device.
24. Services should be available pan Rajasthan including Rural & Remote areas.
25. Device should have Data storage and transfer capability and should have memory capacity: 60 memories for each user X 2 Users and also guest mode should be available.
26. Device should have Arrhythmia detection (APR) & Atrial Fibrillation Detection (AFib) facility.
27. Device should have premature contraction detection.

All other terms & conditions remains the same.

Executive Director (EPM)
RMSCL, Jaipur

Note:- Please note that all above amendments/corrigendum is technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Signature valid

RajKaj Ref
6008950

Digitally signed by Dr. Akash Alha
Designation: Executive Director
Date: 2024.03.07 10:14:55 IST
Reason: Approver