

**Rajasthan Medical Services Corporation Limited, Jaipur**

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmssc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmssc.health.rajasthan.gov.in

No. F-8( ) RMSC/EPM/M-3/18-19/NIB-490/ 4557

Dated: 07/08/2020

**OFFICE ORDER**

The following bidders are declared responsive/non responsive on the basis of documentary & technical evaluation of proposal submitted in response to NIB No. F-8( )RMSC/EPM/M-3/2019-20/NIB-490/4299 dated 13.07.2020, (technical bid opening date 29.07.2020) as per the decision taken in the purchase committee dated 06.08.2020 headed by M.D., RMSCL:-

S. No.	Name of Firm	Quoted items	Remarks	Responsive /Non responsive
1	M/S BPL Medical Technologies Pvt. Ltd. A-654, Ground Floor, Royal Paradise Calgiri Hospital Road, OPP Malviya Nagar Police Station Malviya Nagar Jaipur-302017	<b>Pulse oxymeter (FingerTip)</b>	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically approved किया गया। <b>(All four parameters are autorotating on rotation of pulse oxymeter and visible clearly)</b>	<b>Responsive</b>
2	M/S Microgene Diagnostic Systems Pvt Ltd. 806, Vikram Tower, Rajendra Place, New Delhi - 110008	<b>Pulse oxymeter (FingerTip)</b>	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically not approved किया गया। <b>(All four parameters are not autorotating on rotation of pulse oxymeter )</b>	<b>Non-Responsive</b>
3	M/S Heidelco Medicore Pvt. Ltd. D-1/130 Ground floor, Vibhav Khand, Gomti nagar, Lucknow-226010, UP, India Email: info@heidelco.com	<b>Pulse oxymeter (FingerTip)</b>	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically not approved किया गया। <b>(All four parameters are not autorotating on rotation of pulse oxymeter )</b>	<b>Non-Responsive</b>
4	M/S S.S. Medical Systems Pvt. Ltd. #F2 Bhimtal Industrial Estate, Bhimtal Disstt. Nainital, Uttrakhand-263136 Email: mamta@ssmedworld.com	<b>Pulse oxymeter (FingerTip)</b>	तकनीकी परीक्षण के दौरान फर्म द्वारा प्रस्तुत सेम्पल को तकनीकी समिति द्वारा Technically not approved किया गया। <b>(All four parameters are not autorotating on rotation of pulse oxymeter )</b>	<b>Non-Responsive</b>
5	M/s Medicare Products Inc., C-53A Mansarover Garden, New Delhi-110015	<b>Pulse oxymeter (FingerTip)</b>	तकनीकी मूल्यांकन किये जाने पर फर्म की निविदा आईटम के लिये दस्तावेजी रूप से अयोग्य पाई गई।	<b>Non-Responsive</b>

*(Handwritten signature and date)*  
07/08/2020



			<b>(-Firm did not submit sample along with technical bid submission - Bid security should submit in form of DD/Cheque whereas it providd in form of BG)</b>	
6	M/s Recorder and Medicare Systems Pvt. Ltd., Plot NO. 11. Information Technology Park, HSIIDC, Sector-22, Panchkula 134114 (Haryana)	<b>Pulse oxymeter (FingerTip)</b>	तकनीकी मूल्यांकन किये जाने पर फर्म की निविदा आईटम के लिये दस्तावेजी रूप से अयोग्य पाई गई। <b>{Import Export Certificate (IEC) did not submit in technical bid. It is mandatory document}</b>	<b>Non-Responsive</b>
7	M/s Hello Mobiles, Pvt Ltd	<b>Pulse oxymeter (FingerTip)</b>	तकनीकी मूल्यांकन किये जाने पर फर्म की निविदा आईटम के लिये दस्तावेजी रूप से अयोग्य पाई गई। <b>(- Firm did not provide any Bid submission documents i.e EMD, tender processing fee etc. -Firm also did not submit sample along with technical bid submission)</b>	<b>Non-Responsive</b>

Online financial bids of responsive bidders for above items will be opened on 07.08.2020 at 01:00 p.m.

Executive Director (EPM)  
RMSCL, Jaipur