



**Ministry of Health & Family Welfare
Government of Rajasthan
RMSCL**

“Mukhyamantri Nishulak Dava Yojana”

Gandhi Block, Swasthya Bhawan, Tilak Marg, Jaipur – 302005, India

Phone No: 0141-2228066 , 2228064 Fax No. 0141-2228065 Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067 E_mail : edprocurement@gmail.com, and rmsc.health.rajasthan.gov.in

Ref. No.: F.02(281)/RMSC/Procurement/Item-CV/06/2020/818

Dated:- 28.4.2020

COVID-19

Notice inviting short-term bids for procurement of PPE kit and Three Layer mask.

Rajasthan Medical Services Corporation Limited intends to procure following items as per the quantities given below-

S. N.	Name of Items	Estimated Qty.
1	PPE kit	500000
2	Three Layer Mask	2000000

Interested manufacturers / distributors / dealers / suppliers are requested to send their detailed proposal / offers through e-mail on the following e-mail id – edprocurement@gmail.com contact no. 9829170424 latest by 02nd May 2020 09:00 AM.

Detailed guidelines and specifications may be seen on the website of RMSC which is www.rmschealth.rajasthan.gov.in.

**Managing Director
RMSCL**

Specifications of PPE (Personal Protection Equipments) Kits

1. PPE Kits

1.1 Gloves

- Nitrile
- Non-sterile
- Powder Free
- Outer gloves preferably reach mid-forearm (minimum 280 mm total length)
- Different Sizes (6.5 & 7)
- Quality complaint with the below standards, or equivalent:
 - a. EU Standard directive 93/42/EEC Class I, EN 455
 - b. EU Standard directive 89/686/EEC Category III, EN 374
 - c. ANSI/SEA 105-2011
 - d. ASTM D6319-10

1.2 Coverall (medium and large)*

- Impermeable to blood and body fluids
- Single use
- Avoid culturally unacceptable colors e.g. black
- Light colors are preferable to better detect possible contamination
- Thumb/finger loops to anchor sleeves in place
- Quality complaint with following standard
 - a. Meets or exceeds ISO 16603 class 3 exposure pressure, or equivalent
- Coverall suit/ Jump Suit-Earlier SPEC was minimum 60GSM and now its revised to 95GSM- splash proof/ Non permeable quality CE STD. 02091

1.3 Goggles

- With transparent glasses, zero power, well fitting, covered from all sides with elastic band/ or adjustable holder
- Good seal with skin of the face
- Flexible frame to easily fir all faces contours without too much pressure
- Covers the eyes and the surrounding areas and accommodates for prescription glasses

- Fog and scratch resistant
- Adjustable band to secure firmly so as not to become loose during clinical activity
- Indirect venting to reduce fogging
- May be re-usable (provided appropriate arrangements for decontamination are in place) or disposable
- Quality compliant with the below standards, or equivalent
 - a. EU standard directive 86/686/EEC, EN 166/2002
 - b. ANSI/SEA Z87.1 – 2010

1.4 Shoe Covers

- Made up of the same fabric as of coverall
- Should cover the entire shoe and reach above ankles

1.5 Face Shield

- Made of clear plastic and provides good visibility to both the wearer and the patient
- Adjustable band to attach firmly around the head and fit snugly against the forehead
- Fog resistant (preferable)
- Completely covers the sides and length of the face
- May be re-usable (made of material which can be cleaned and disinfected) or disposable
- Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 86/686/EEC, EN 166/2002
 - b. ANSI/SEA Z87.1 -2010

Specifications Triple Layer Medical Mask

- Three layered medical mask of non – woven material with nose piece, having filter efficiency of 99% for 3 micron particle size.
 - a. ISI Specification or equivalent.

Undertaking

Undertake that since management of COVID-19 is an emergency, RMSCL will issue orders for supply anytime after 2nd May 2020 09:00AM to the best price available keeping in mind earliest supply in quantities which the RMSCL thinks fit. The decision of RMSCL shall be final and not subject to any claim or dispute.

Name of firm

Authorized Representative

Signature & Date