

Rajasthan Medical Services Corporation Limited
Gandhi Block, SwasthyaBhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 ,2228064

Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E_mail : edpssrmisc@rajasthan.gov.in

Ref. No.:- F.02(130)/RMSCL/Proc./S&S(MD)/NIB-02/2023/ 1287 Dated:- 16/02/2023

Corrigendum – I

Subject:- Deletion and addition of items from Annexure VII and BoQ

Ref:- NIB No. F.02(130)/RMSCL/Proc./S&S(MD)/NIB-02/2023/1192 Dated: 07.02.2023
(Technical bid opening due on dated –28.02.2023)

The following amendments are made:

S. No	Existing Annexure VII, BoQ		Amended Annexure VII, BoQ
1.	Annexure VII, BoQ		Annexure VII, BoQ
	S. No.	Code No.	Name of approved item (s) with specification
	1	S-26	Sterile Hypodermic Syringe with Needle/or with Needle attached , 24G, Single Use - 2 ml <ul style="list-style-type: none"> • Clear transparent chamber • Prominent graduation • Inert material gasket at the piston to minimize friction during movement & prevent leakage and back flow • Sharp needle ensuring minimum trauma during penetration • Shall conform to relevant IS • Packing: Needle should be with the syringe and packed in unit ribbon pack • The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.
	2	S-27	Sterile Hypodermic Syringe with Needle/or with Needle attached , 24G, Single Use - 5 ml <ul style="list-style-type: none"> • Clear transparent chamber • Prominent graduation • Inert material gasket at the piston to minimize friction during movement & prevent leakage and back flow • Sharp needle ensuring minimum trauma during penetration • Shall conform to relevant IS • Packing: Needle should be with the syringe and packed in unit ribbon pack • The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.
	3	S-28	Sterile Hypodermic Syringe with Needle/or with Needle attached , 22G, Single Use - 10 ml <ul style="list-style-type: none"> • Clear transparent chamber • Prominent graduation • Inert material gasket at the piston to minimize friction during
	Items Deleted:-		
	S. No.	Code No.	
	1	S-26	deleted
	2	S-27	deleted
	3	S-28	deleted
	4	S-29	deleted

		<p>movement & prevent leakage and back flow</p> <ul style="list-style-type: none"> • Sharp needle ensuring minimum trauma during penetration • Shall conform to relevant IS • Packing: Needle should be with the syringe and packed in unit ribbon pack • The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.
	S-29	<p>Sterile Hypodermic Syringe with Needle/or with Needle attached , 22G, Single Use - 20 ml</p> <ul style="list-style-type: none"> • Clear transparent chamber • Prominent graduation • Inert material gasket at the piston to minimize friction during movement & prevent leakage and back flow • Sharp needle ensuring minimum trauma during penetration • Shall conform to relevant IS • Packing: Needle should be with the syringe and packed in unit ribbon pack • The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.

2. **Items Added in Annexure VII, BoQ**

S. No	Code No.	Name of approved item (s) with specification	Minimum labelled Shelf life (In months)	Packing Unit	Class As per Medical Device Rules 2017	Tender quantity
1	S-26 (a)	<p>Sterile Hypodermic Syringe with Needle, 24G, Single Use - 2 ml</p> <ul style="list-style-type: none"> • Clear transparent chamber • Prominent graduation • Inert material gasket at the piston to minimise friction during movement & prevent leakage and back flow • Sharp needle ensuring minimum trauma during penetration • Shall conform to relevant IS for syringe - IS 10258 and for needle IS 10654 • Packing: Needle should be with the syringe and packed in unit pack • The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container 	36 months	Unit	Class B	81826862
2	S-27(a)	<p>Sterile Hypodermic Syringe with Needle, 24G, Single Use - 5 ml</p> <ul style="list-style-type: none"> • Clear transparent chamber • Prominent graduation • Inert material gasket at the piston to minimise friction during movement & prevent leakage and back flow • Sharp needle ensuring minimum trauma during penetration • Shall conform to relevant IS for syringe - IS 10258 and for needle IS 10654 	36 months	Unit	Class B	128162684

		<ul style="list-style-type: none"> • Packing: Needle should be with the syringe and packed in unit pack • The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container. 				
3	S-28(a)	<ul style="list-style-type: none"> • Sterile Hypodermic Syringe with Needle, 22G, Single Use - 10 ml • Clear transparent chamber • Prominent graduation • Inert material gasket at the piston to minimise friction during movement & prevent leakage and back flow • Sharp needle ensuring minimum trauma during penetration • Shall conform to relevant IS for syringe - IS 10258 and for needle IS 10654 • Packing: Needle should be with the syringe and packed in unit pack • The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container. 	36 months	Unit	Class B	75208056
4	S-29(a)	<ul style="list-style-type: none"> • Sterile Hypodermic Syringe with Needle, 22G, Single Use - 20 ml • Clear transparent chamber • Prominent graduation • Inert material gasket at the piston to minimise friction during movement & prevent leakage and back flow • Sharp needle ensuring minimum trauma during penetration • Shall conform to relevant IS for syringe - IS 10258 and for needle IS 10654 • Packing: Needle should be with the syringe and packed in unit pack • The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container. 	36 months	Unit	Class B	4771104

Note:-

- It may be noted that if any further amendments are issued then a corrigendum will be published and informed.
- Rest of the terms and conditions will remain the same.

**Executive Director (Proc.)
RMSCL**

